



Certificate of Compliance

Certificate: 1008476

Master Contract: 216815

Project: 2548740

Date Issued: August 16, 2012

Issued to: BriskHeat Corp

1055 Gibbard Ave
Columbus, OH 43201
USA

Attention: Jason Adams

The products listed below are eligible to bear the CSA Mark shown



Simon Lee

Issued by: Simon Lee

PRODUCTS

CLASS 2841 01 - HEATERS - Industrial and Laboratory

Laboratory Heating Mantles rated up to 15A as follows:

Model No.(120Vac)	Model No.(240Vac)	Total Watts
HM0050-HS1 or THM0050-HS1	HM0050-HS2 or THM0050-HS2	60
HM0100-HS1 or THM0100-HS1	HM0100-HS2 or THM0100-HS2	80
HM0125-HS1 or THM0125-HS1	HM0125-HS2 or THM0125-HS2	80
HM0200-HS1 or THM0200-HS1	HM0200-HS2 or THM0200-HS2	100



Certificate: 1008476

Master Contract: 216815

Project: 2548740

Date Issued: August 16, 2012

HM0250-HS1 or THM0250-HS1 or HM0250-TS1 or THM0250-TS1	HM0250-HS2 or THM0250-HS2 or HM0250-TS2 or THM0250-TS2	180
HM0500-HS1 or THM0500-HS1 or HM0500-TS1 or THM0500-TS1	HM0500-HS2 or THM0500-HS2 or HM0500-TS2 or THM0500-TS2 or HM0500-SS2 or THM0500-SS2	270
HM1000-HS1 or THM1000-HS1 or HM1000-TS1 or THM1000-TS1	HM1000-HS2 or THM1000-HS2 or HM1000-TS2 or THM1000-TS2 or HM1000-SS2 or THM1000-SS2	380
HM2000-HS1 or THM2000-HS1 or HM2000-TS1 or THM2000-TS1	HM2000-HS2 or THM2000-HS2 or HM2000-TS2 or THM2000-TS2 or HM2000-SS2 or THM2000-SS2	500
HM3000-HS1 or THM3000-HS1 or HM3000-TS1 or THM3000-TS1	HM3000-HS2 or THM3000-HS2 or HM3000-TS2 or THM3000-TS2	500
HM5000-HS1 or THM5000-HS1 or HM5000-TS1 or THM5000-TS1	HM5000-HS2 or THM5000-HS2 or HM5000-TS2 or THM5000-TS2	600
HM12000-HS1 or THM12000-HS1	HM12000-HS2 or THM12000-HS2	1300

Model No.(120Vac)	Model No.(240Vac)	Total Watts
HM0050VF1	HM0050VF2	52
HM0100VF1	HM0100VF2	70
HM0250VF1	HM0250VF2	143
HM0500VF1	HM0500VF2	210
HM1000VF1	HM1000VF2	377
HM2000VF1	HM2000VF2	520
HM3000VF1	HM3000VF2	550



Certificate: 1008476

Master Contract: 216815

Project: 2548740

Date Issued: August 16, 2012

HM5000VF1	HM5000VF2	620
-----------	-----------	-----

Note: The heating mantles are Certified only as component devices for use in other equipment where the suitability of the combination is to be determined by CSA International.

APPLICABLE REQUIREMENTS

CSA Std. C22.2 No. 88-1958 - Industrial Heating Equipment